

	<b>QUALITY STANDARD</b> s.r.l. Via dei Chiosi,2 20040 - Cavenago Brienza - MI	<b>RESISTANCE TO DUST, SOLID OBJECTS AND MOISTURE</b>  CEI EN 60529 II ED.	Rif. n° : 99/RPR/105  All. n°1 Rev.0 Pag. 1 di 1
--	--	--	--

Mark: REVALCO S.r.l.	Type: 72x72
Object: Panel instrument	Screws Locking (Torque) : / Nm

DUST TEST	
Dust quantity : 2 Kg/m <sup>3</sup>	Time test : 8h
Box dimension : 4.5m <sup>3</sup>	Depression into the box : /

WATER TEST	
Diameter water lance : /	Time test : 10min.
Range water : 0.84l/min±5%	Distance: 200mm

- classification according to IP .....	54	-
- fixing screws tightened; torque (Nm)...	/	-
- tests according to clauses .....	13.4-13.5-14.2.4	-
- Position test	Normal position	-
- electric strength		P
1) no deposit in dust-proof into the box		P
2) no talcum in dust-tight into the box		NA
3) no trace of water on live parts		P
4) no accumulation of water in waterproof box		P
5) no water in watertight box		NA
6) no contact with live parts		P

Possible test case verdicts:

- test case does not apply to the test object .....: N(.A.)
- test object does meet the requirement.....: P(ass)
- test object does not meet the requirement .....: F(ail)